



CENTRAL POLLUTION CONTROL BOARD
HEAD OFFICE - DELHI
Parivesh Bhavan, East Arjun Nagar, Delhi - 110032



TC-7723

Trace Organics Laboratory
Organochlorine Pesticides Analysis Report

ULR No: TC772323400000015F

1.	Report No. & Date of Issue	:	TOL/2324/OCPS/00015,19/10/2023
2.	Report sent to (Name,Mobile no. & Address of Indentor)	:	DH of Instrumentation Lab
3.	Samples Registration No. & Date	:	TOL/OCPS/2324/SR00011,22/09/2023
4.	Analysis Request Division/Organization	:	Instrumentation Lab
5.	Sample Collected by	:	Yogita Kharayat,Dr. L. N. Gupta and Sh. Ashutosh Kumar
6.	Sampling Plan Preference	:	
7.	Sample Sealing Status	:	Not-Sealed
8.	Date & Time of Sample Receipt	:	22/09/2023 15:25 PM
9.	Sample Analysis Period	:	25/09/2023 27/09/2023
10.	Sample Details	:	Drinking water, CPCB, HO, Delhi
11.	Report Status	:	Final

Sr. No	Field Code	Parameter Name	Date of Sample Collection	Sample Matrix	Unit	Result value
1	Ground Floor, CPCB, Dekli	Aldrin	21/09/2023	Fresh Water	O	BDL
2	Ground Floor, CPCB, Dekli	Dieldrin	21/09/2023	Fresh Water		BDL
3	Ground Floor, CPCB, Dekli	o,p DDT	21/09/2023	Fresh Water		BDL
4	Ground Floor, CPCB, Dekli	p,p DDE	21/09/2023	Fresh Water		BDL
5	Ground Floor, CPCB, Dekli	p,p DDT	21/09/2023	Fresh Water		BDL
6	Ground Floor, CPCB, Dekli	-	21/09/2023	Fresh Water		BDL
7	Ground Floor, CPCB, Dekli	=#=#	21/09/2023	Fresh Water		BDL
8	Ground Floor, CPCB, Dekli	Endosulfan	21/09/2023	Fresh Water	O	BDL
9	Ground Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water	O	BDL
10	Ground Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water		BDL
11	Ground Floor, CPCB, Dekli	p,p DDD	21/09/2023	Fresh Water		BDL
12	1st Floor, CPCB, Dekli	Aldrin	21/09/2023	Fresh Water	O	BDL
13	1st Floor, CPCB, Dekli	Dieldrin	21/09/2023	Fresh Water		BDL
14	1st Floor, CPCB, Dekli	o,p DDT	21/09/2023	Fresh Water		BDL
15	1st Floor, CPCB, Dekli	p,p DDE	21/09/2023	Fresh Water		BDL
16	1st Floor, CPCB, Dekli	p,p DDT	21/09/2023	Fresh Water		BDL



CENTRAL POLLUTION CONTROL BOARD
HEAD OFFICE - DELHI
Parivesh Bhavan, East Arjun Nagar, Delhi - 110032



Trace Organics Laboratory
Organochlorine Pesticides Analysis Report

ULR No: TC772323400000015F

17	1st Floor, CPCB, Dekli	-	21/09/2023	Fresh Water		BDL
18	1st Floor, CPCB, Dekli	=#=#	21/09/2023	Fresh Water		BDL
19	1st Floor, CPCB, Dekli	Endosulfan	21/09/2023	Fresh Water	O	BDL
20	1st Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water	O	BDL
21	1st Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water		BDL
22	1st Floor, CPCB, Dekli	p,p DDD	21/09/2023	Fresh Water		BDL
23	2nd Floor, CPCB, Dekli	Aldrin	21/09/2023	Fresh Water	O	BDL
24	2nd Floor, CPCB, Dekli	Dieldrin	21/09/2023	Fresh Water		BDL
25	2nd Floor, CPCB, Dekli	o,p DDT	21/09/2023	Fresh Water		BDL
26	2nd Floor, CPCB, Dekli	p,p DDE	21/09/2023	Fresh Water		BDL
27	2nd Floor, CPCB, Dekli	p,p DDT	21/09/2023	Fresh Water		BDL
28	2nd Floor, CPCB, Dekli	-	21/09/2023	Fresh Water		BDL
29	2nd Floor, CPCB, Dekli	=#=#	21/09/2023	Fresh Water		BDL
30	2nd Floor, CPCB, Dekli	Endosulfan	21/09/2023	Fresh Water	O	BDL
31	2nd Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water	O	BDL
32	2nd Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water		BDL
33	2nd Floor, CPCB, Dekli	p,p DDD	21/09/2023	Fresh Water		BDL
34	3rd Floor, CPCB, Dekli	Aldrin	21/09/2023	Fresh Water	O	BDL
35	3rd Floor, CPCB, Dekli	Dieldrin	21/09/2023	Fresh Water		BDL
36	3rd Floor, CPCB, Dekli	o,p DDT	21/09/2023	Fresh Water		BDL
37	3rd Floor, CPCB, Dekli	p,p DDE	21/09/2023	Fresh Water		BDL
38	3rd Floor, CPCB, Dekli	p,p DDT	21/09/2023	Fresh Water		BDL
39	3rd Floor, CPCB, Dekli	-	21/09/2023	Fresh Water		BDL
40	3rd Floor, CPCB, Dekli	=#=#	21/09/2023	Fresh Water		BDL
41	3rd Floor, CPCB, Dekli	Endosulfan	21/09/2023	Fresh Water	O	BDL



CENTRAL POLLUTION CONTROL BOARD
HEAD OFFICE - DELHI
Parivesh Bhavan, East Arjun Nagar, Delhi - 110032



Trace Organics Laboratory
Organochlorine Pesticides Analysis Report

ULR No: TC772323400000015F

42	3rd Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water	O	BDL
43	3rd Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water		BDL
44	3rd Floor, CPCB, Dekli	p,p DDD	21/09/2023	Fresh Water		BDL
45	4th Floor, CPCB, Dekli	Aldrin	21/09/2023	Fresh Water	O	BDL
46	4th Floor, CPCB, Dekli	Dieldrin	21/09/2023	Fresh Water		BDL
47	4th Floor, CPCB, Dekli	o,p DDT	21/09/2023	Fresh Water		BDL
48	4th Floor, CPCB, Dekli	p,p DDE	21/09/2023	Fresh Water		BDL
49	4th Floor, CPCB, Dekli	p,p DDT	21/09/2023	Fresh Water		BDL
50	4th Floor, CPCB, Dekli	-	21/09/2023	Fresh Water		BDL
51	4th Floor, CPCB, Dekli	=#=#	21/09/2023	Fresh Water		BDL
52	4th Floor, CPCB, Dekli	Endosulfan	21/09/2023	Fresh Water	O	BDL
53	4th Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water	O	BDL
54	4th Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water		BDL
55	4th Floor, CPCB, Dekli	p,p DDD	21/09/2023	Fresh Water		BDL
56	5th Floor, CPCB, Dekli	Aldrin	21/09/2023	Fresh Water	O	BDL
57	5th Floor, CPCB, Dekli	Dieldrin	21/09/2023	Fresh Water		BDL
58	5th Floor, CPCB, Dekli	o,p DDT	21/09/2023	Fresh Water		BDL
59	5th Floor, CPCB, Dekli	p,p DDE	21/09/2023	Fresh Water		BDL
60	5th Floor, CPCB, Dekli	p,p DDT	21/09/2023	Fresh Water		BDL
61	5th Floor, CPCB, Dekli	-	21/09/2023	Fresh Water		BDL
62	5th Floor, CPCB, Dekli	=#=#	21/09/2023	Fresh Water		BDL
63	5th Floor, CPCB, Dekli	Endosulfan	21/09/2023	Fresh Water	O	BDL
64	5th Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water	O	BDL
65	5th Floor, CPCB, Dekli	HCH	21/09/2023	Fresh Water		BDL
66	5th Floor, CPCB, Dekli	p,p DDD	21/09/2023	Fresh Water		BDL



CENTRAL POLLUTION CONTROL BOARD
HEAD OFFICE - DELHI
Parivesh Bhavan, East Arjun Nagar, Delhi - 110032



Trace Organics Laboratory
Organochlorine Pesticides Analysis Report

ULR No: TC772323400000015F

BDL : Below Deduction Limit

Remarks (if any) :

Statement :

1. The results relate only to the samples tested.
2. The report shall not be reproduced except in full, without the written approval of the laboratory.

Analyst

Supervisor & Reviewer

Issued By(DH TOL-Lab)

Charu Sharma

Bhupander Kumar

Sanjay Kumar

Doc: CB/CL/QR/7.8/TOL-1	Issue No.:01	Revision No : 08	Issue Date: 18/12/2017	Revision Date: 05/05/2022
-------------------------	--------------	------------------	------------------------	---------------------------

Date: 02 Nov 2023

This is a computer-generated Report. No signature is required.

Parameter Name	Test Method	Unit	Limit of Detection	Uncertainty of Measurement
Aldrin	USEPA Method No. 8081B, 2007	O	0.010	. O . . O
Dieldrin	USEPA Method No. 8081B, 2007		0.010	. O . . O
o,p DDT	USEPA Method No. 8081B, 2007		0.010	. O . . O
p,p DDD	USEPA Method No. 8081B, 2007		0.010	. O . . O
p,p DDE	USEPA Method No. 8081B, 2007		0.010	. O . . O
p,p DDT	USEPA Method No. 8081B, 2007		0.010	. O . . O
-	USEPA Method No. 8081B, 2007		0.010	. O . . O
=#=#	USEPA Method No. 8081B, 2007		0.010	. O . . O
Endosulfan	USEPA Method 8081B : 2007	O	0.010	. O . . O
HCH	USEPA Method No. 8081B, 2007	O	0.010	. O . . O
HCH	USEPA Method No. 8081B, 2007		0.010	. O . . O