



Form No.: QRF/2022-27/Rev-5

CALIBRATION CERTIFICATE OF : ELECTRONIC BALANCE

PAGE 01 OF 5

- 1.0 SCOPE
- 1.1 Service Request No. : CA/0021/11-22
- 1.2 Calibration Certification No. : ERTL(E)/CAL/C158/0026/11-22 /NABL
Date : 28-11-2022
- ULR No. : CC200822000001238F
- 1.3 Requested by : CENTRAL POLLUTION CONTROL BOARD
(Name & address of the organisation) 502 SOUTHEND CONCLAVE, 6TH FL..1582 RAJDANGA MAIN ROAD
KOLKATA-700107
WEST BENGAL
- 1.4 Description of the item to be calibrated
- | | |
|--------|----------------------|
| Item | : ELECTRONIC BALANCE |
| Make | : CONTECH |
| Model | : CAS-164 |
| sl.No. | : 1565971 |
| Id No. | : NIL |
- 1.4.1 Applicable Specification : Not mentioned
- 1.4.2 Characterisation and condition of item
- | | |
|------------------|-------------------|
| Characterisation | : Not applicable/ |
| Condition | : Satisfactory |
- 1.5 Date of receipt of item : 23-11-2022
- 1.6 Date of start of calibration : 23-11-2022
- 1.6.1 Date of completion of calibration : 23-11-2022
- 1.7 Location where calibration performed : ON SITE (Refer section 3.0 for address)
- 1.8 Ambient condition during measurement : 27+/-5 °C
45-70 %RH
- 1.9 Discipline (NABL) : OPTICAL DISCIPLINE
- 1.9.1 Calibration procedure followed : EP15/OPN062
- 1.9.2 Details of non standard method followed (if any):
- 1.10 Measurement Uncertainty : The uncertainty figure is stated in the result.
- 1.11 Traceability : NPL,NEW DELHI ,

The following Standards / Equipment have been used .

EQPT_NO	ITEM	MAKE	MODEL	CAL.VALID UPTO
1366	STANDARD WEIGHT SET	SARTORIUS AG	YCS011-612-02	26-10-2024

Note: All calibration were conducted within the validity period of respective equipment shown above.



CALIBRATION CERTIFICATE NO :				ERTL(E)/CAL/C158/0026/11-22/NABL	
CALIBRATION CERTIFICATE OF:				DATE	CALIBRATION ENGINEER
ELECTRONIC BALANCE				23-11-2022	P.DHABAL RAVI RANJAN
TYPE : SINGLE PAN	MAKE:	CONTECH	MODEL:	CAS-164	
JOB NO:	0026/CA/11-22	SL NO :	1565971	ID NO :	NIL
Range:	upto 160 g	Readability :	0.0001 g	Accuracy :	Not Specified
TEST CONDITION:	Amb. Temp. 25.6	Relative Humidity 57 % R.H.	Location BALANCE ROOM		
Measurement Results	MECHANICAL DISCIPLINE				

(a) Linearity Test : (Varified against Calibrated Mass)

The measurement of indication error (i.e. departure of displayed reading from nominal Value) at three points evenly spread across the Calibration range.

Cal. Point	Standard Mass value in g	Balance Reading in g	Expanded Uncertainty at 95% CL & k=2 in mg.(+/-)
1	0.002000	0.0018	0.10
2	0.004999	0.0051	0.10
3	0.010004	0.0101	0.10
4	0.020003	0.0199	0.10
5	0.050004	0.0500	0.10
6	0.100007	0.1001	0.10
7	0.200007	0.2001	0.10
8	0.500009	0.5002	0.10
9	0.999994	0.9999	0.10
10	2.000003	2.0002	0.10
11	5.000004	5.0004	0.10
12	10.000012	10.0015	0.10
13	20.000021	20.0034	0.10
14	50.000010	50.0076	0.10
15	99.999920	100.0168	0.10
16	149.999930	150.0261	0.12
17	154.999934	155.0265	0.16

Thermal Stabilisation Hour : 2

Calibration Certificate issued for balance shall be used for Scientific or Industrial purpose only

Authorised Signatory :
Date : 25/11/22
Paresh Dhabal
Scientist 'C'
STQC, ERTL (E)
Govt. of India, Min. of Electronics & IT
Block-DN, Sector-V, Salt Lake, Kolkata-91



CALIBRATION CERTIFICATE NO :			ERTL(E)/CAL/C158/0026/11-22/NABL		
CALIBRATION CERTIFICATE OF:			DATE	CALIBRATION ENGINEER	
ELECTRONIC BALANCE			23-11-2022	P.DHABAL RAVI RANJAN NIL	
JOB NO:	0026/CA/11-22	SL NO :	CAS-164	ID NO :	
			1565971		
Range:	Readability :				Accuracy :
upto 160 g	'0.0001 g				Not Specified
TEST CONDITION:	Amb. Temp.	Humidity	Location		
	25.6	57 % R.H.	BALANCE ROOM		
Measurement Results	MECHANICAL DISCIPLINE				

(b) Repeatability Test : (Verified against Calibrated Mass)
The Measurement of the ability to display the same result when repeated measurement are made under the same weighing Condition.

Cal. Point	Std. Mass Used in g	Unloaded Reading in g	Loaded Reading in g	Max Difference in g	Expanded Uncertainty in \pm mg at $k=2$ & 95% C.L
1	99.99992	0.0000	100.0168	0.0004	0.10
2		0.0000	100.0167		
3		0.0000	100.0168		
4		0.0001	100.0165		
5		0.0000	100.0168		
6		0.0001	100.0167		
7		0.0000	100.0168		
8		0.0000	100.0166		
9		0.0000	100.0167		
10		0.0001	100.0166		

(c) Eccentricity Test : (Verified against Calibrated Mass)
The measurement of the indicated error when the load is applied off centre.
The mass position as per figure 1.

Load Position	Standard Value in g	Balance Reading in g	Max Eccentricity Error in g	Expanded Uncertainty in \pm mg at $k=2$ & 95% C.L
1	99.99992	100.0167	0.0005	0.27
2		100.0166		
3		100.0165		
4		100.0170		
5		100.0172		
6		100.0168		

Thermal Stabilisation Hour : 2

Calibration Certificate issued for balance shall be used for Scientific or Industrial purpose only




Figure: 1

Authorised Signatory :
Date : 25/11/22
Paresh Dhabal
Scientist 'C'
STQC, ERTL (E)
Govt. of India, Min. of Electronics & IT
Block-DN, Sector-V, Salt Lake, Kolkata-91



Form No.: QRF/2022-27/Rev-5

CALIBRATION CERTIFICATE OF : ELECTRONIC BALANCE

PAGE 4 OF 5

Calibration Certification No. : ERTL(E)/CAL/C158/0026/11-22 /NABL
Date : 28-11-2022
ULR No. : CC200822000001238F

3.0 Remarks (if any) :

Performance check has been carried out.

ON SITE Address

Same as 1.3



4.0 Ref. of sticker allotted : 03421

5.0 Next Calibration due date : 23-11-2023



Scan to verify the originality of the report



CC2008



Scan to verify NABL certificate



Form No.: QRF/2022-27/Rev-5

CALIBRATION CERTIFICATE OF : ELECTRONIC BALANCE

PAGE 5 OF 5

Calibration Certification No. : ERTL(E)/CAL/C158/0026/11-22 /NABL
Date : 28-11-2022

ULR No. : CC200822000001238F

NOTES

1. This certificate/ report refer only to the particular item submitted for calibration/test.
2. The certificate/report, if reproduced for any purpose, commercial or otherwise, should be reproduced in full. The reproduction of a part, or an abstract thereof, has to be got specially approved from Director, Electronics Regional Test Laboratory (East).
3. The results reported in the certificate /reports are valid at the time of and under the stated conditions on measurement.
4. Calibration need to be done periodically to maintain the accuracy of the measurement. (in case of calibration report only)

R. Bhattacharya
RELEASED BY
01/12/2022
(Signature & Date)

रत्ना भट्टाचार्य Ratna Bhattacharyya
वैज्ञानिक 'ई' Scientist 'E'
भारत सरकार Government of India
ई आर टी एल (ई) ERTL (East)
सॉल्ट लेक, कोलकाता-700091 Salt Lake, Kolkata-700091