



Form No.: QRF/2022-27/Rev-5

CALIBRATION CERTIFICATE OF : ELECTRONIC BALANCE

PAGE 01 OF 5

- 1.0 SCOPE
- 1.1 Service Request No. : CA/0021/11-22
- 1.2 Calibration Certification No. : ERTL(E)/CAL/C158/0025/11-22 /NABL
Date : 28-11-2022
- ULR No. : CC200822000001237F
- 1.3 Requested by : CENTRAL POLLUTION CONTROL BOARD
(Name & address : 502 SOUTHEND CONCLAVE, 6TH FL..1582 RAJDANGA MAIN
ROAD
of the organisation) KOLKATA-700107
WEST BENGAL
- 1.4 Description Item : ELECTRONIC BALANCE
Identification Make : ACZET
of the item Model : CX-265N
to be calibrated sl.No. : 133-04A
Id No. : NIL
- 1.4.1 Applicable Specification : Not mentioned
- 1.4.2 Characterisation and Condition : Not applicable/
condition of item : Satisfactory
- 1.5 Date of receipt of item : 23-11-2022
- 1.6 Date of start of calibration : 23-11-2022
- 1.6.1 Date of completion of calibration : 23-11-2022
- 1.7 Location where calibration performed : ON SITE (Refer section 3.0 for address)
- 1.8 Ambient condition during measurement : 27+/-5 °C
45-70 %RH
- 1.9 Discipline (NABL) : OPTICAL DISCIPLINE
- 1.9.1 Calibration procedure followed : EP15/OPN062
- 1.9.2 Details of non standard method followed (if any):
- 1.10 Measurement Uncertainty : The uncertainty figure is stated in the result.
- 1.11 Traceability : NPL,NEW DELHI ,

The following Standards / Equipment have been used .

EQPT_NO	ITEM	MAKE	MODEL	CAL.VALID UPTO
1366	STANDARD WEIGHT SET	SARTORIUS AG	YCS011-612-02	26-10-2024

Note: All calibration were conducted within the validity period of respective equipment shown above.




CALIBRATION CERTIFICATE NO :			ERTL(E)/CAL/C158/0025/11-22/NABL		
CALIBRATION CERTIFICATE OF:			DATE	CALIBRATION ENGINEER	
ELECTRONIC BALANCE			23-11-2022	P.DHABAL RAVI RANJAN NIL	
MAKE: ACZET		MODEL: CX-265 N			
JOB NO:	0025/CA/11-22	SL NO :	133-04A	ID NO :	
Range: upto 220 g	Readability : 0.00001/0.0001 g			Accuracy : Not Specified	
TEST CONDITION:	Amb. Temp. 25.7	Relative Humidity 57 % R.H.		Location BALANCE ROOM	
Measurement Results	MECHANICAL DISCIPLINE				

(a) Linearity Test : (Varified against Calibrated Mass)
The measurement of indication error (i.e. depature of displayed reading from nominal Value)
at three points evenly spread across the Calibration range.

Cal. Point	Standard Mass value in g	Balance Reading in g	Expanded Uncertainty at 95% CL & k=2 in mg.(+/-)
1	0.002000	0.00200	0.10
2	0.004999	0.00501	0.10
3	0.010004	0.00994	0.10
4	0.020003	0.01991	0.10
5	0.050004	0.04993	0.10
6	0.100007	0.09988	0.10
7	0.200007	0.19988	0.10
8	0.500009	0.49987	0.10
9	0.999994	0.99986	0.10
10	2.000003	1.99983	0.10
11	5.000004	4.99992	0.10
12	10.000012	9.99995	0.10
13	20.000021	20.00001	0.10
14	50.000010	49.99991	0.10
15	99.999920	99.9997	0.10
16	200.000130	200.0007	0.13
17	220.000151	220.0009	1.50

Thermal Stabilisation Hour : 2

Calibration Certificate issued for balance shall be
used for Scientific or Industrial purpose only

Authorised Signatory : 
Date :
Parash Dhabal
Scientist 'C'
STQC, ERTL (E)
Govt. of India, Min. of Electronics & IT
Block-DN, Sector-V, Salt Lake, Kolkata-91



Page 3 of 5

CALIBRATION CERTIFICATE NO :		ERTL(E)/CAL/C158/0025/11-22/NABL		
CALIBRATION CERTIFICATE OF :		ELECTRONIC BALANCE		CALIBRATION ENGINEER
MAKE: ACZET		MODEL: CX-265 N	DATE: 23-11-2022	P.DHABAL RAVI RANJAN
JOB NO: 0025/CA/11-22	SL NO: 133-04A	ID NO:	NIL	
Range: upto 220 g	Readability: '0.00001/0.0001 g	Accuracy: Not Specified		
TEST CONDITION:	Amb. Temp. 25.7	Humidity 57 % R.H.	Location BALANCE ROOM	
Measurement Results	MECHANICAL DISCIPLINE			

(b) Repeatability Test : (Verified against Calibrated Mass)
The Measurement of the ability to display the same result when repeated measurement are made under the same weighing Condition.

Cal. Point	Std. Mass Used in g	Unloaded Reading in g	Loaded Reading in g	Max Difference in g	Expanded Uncertainty in \pm mg at $k=2$ & 95% C.L
1	99.99992	0.0000	99.9998	0.0003	0.10
2		0.0000	99.9997		
3		0.0001	99.9999		
4		0.0000	99.9996		
5		0.0000	99.9999		
6		0.0000	99.9998		
7		0.0000	99.9997		
8		0.0000	99.9998		
9		0.0000	99.9997		
10		0.0000	99.9998		

(c) Eccentricity Test : (Verified against Calibrated Mass)
The measurement of the indicated error when the load is applied off centre.
The mass position as per figure 1.

Load Position	Standard Value in g	Balance Reading in g	Max Eccentricity Error in g	Expanded Uncertainty in \pm mg at $k=2$ & 95% C.L
1	99.99992	99.9998	0.0004	0.24
2		99.9995		
3		100.0002		
4		99.9997		
5		99.9998		
6		99.9998		

Thermal Stabilisation Hour : 2

Calibration Certificate issued for balance shall be used for Scientific or Industrial purpose only




Figure: 1

Authorised Signatory:
Date: 23/11/22
Paresb Dhabal
Scientist 'C'
STQC, ERTL (E)
Govt. of India, Min. of Electronics & IT
Block-DN, Sector-V, Salt Lake, Kolkata-91



Form No.: QRF/2022-27/Rev-5

CALIBRATION CERTIFICATE OF : ELECTRONIC BALANCE

PAGE 4 OF 5

Calibration Certification No. : ERTL(E)/CAL/C158/0025/11-22 /NABL
Date : 28-11-2022
ULR No. : CC200822000001237F

3.0 Remarks (if any) :

Performance check has been carried out.

ON SITE Address

Same as 1.3



4.0 Ref. of sticker allotted : 03422

5.0 Next Calibration due date : 23-11-2023



Scan to verify the originality
of the report



CC2008



Scan to verify NABL
certificate



Form No.: QRF/2022-27/Rev-5

CALIBRATION CERTIFICATE OF : ELECTRONIC BALANCE

PAGE 5 OF 5

Calibration Certification No. : ERTL(E)/CAL/C158/0025/11-22 /NABL
Date : 28-11-2022

ULR No. : CC200822000001237F

NOTES

1. This certificate/ report refer only to the particular item submitted for calibration/test.
2. The certificate/report, if reproduced for any purpose, commercial or otherwise, should be reproduced in full. The reproduction of a part, or an abstract thereof, has to be got specially approved from Director, Electronics Regional Test Laboratory (East).
3. The results reported in the certificate /reports are valid at the time of and under the stated conditions on measurement.
4. Calibration need to be done periodically to maintain the accuracy of the measurement. (in case of calibration report only)

R. Bhattacharya
RELEASED BY
01/12/2022
(Signature & Date)

रत्ना भट्टाचार्य Ratna Bhattacharyya
वैज्ञानिक 'ई' Scientist 'E'
भारत सरकार Government of India
ई आर टी एल (ई) ERTL (East)
सॉल्ट लेक, कोलकाता-700091 Salt Lake, Kolkata-700091