

Sample Requisition No - RDKL/WW/2324/SR00061

Allocation Date - 30/11/2023

Parameter - pH

Test Start Date - 30/11/2023

Test End Date - 30/11/2023

Due date - 15/12/2023

Name of Supervisor - Md. A. Rafique, Sc-B'

Sl.No.	Sample Code	pH Value	Temp(°C)
01	Std.	pH 4.0 $\approx$ 4.0	25
02	Std.	pH 7.0 $\approx$ 7.0	25
03	Std.	pH 9.18 $\approx$ 9.16	25
04	2324/RDKL/STW/WW/L00132	7.3	25
05	" " /L00133	7.4	25
06	" /L00134	7.3	25

Ajmal Nazki  
02/12/2023

Amal  
07/12/23

Sample Requisition No- RDKL/WW/2324/SR00061

Allocation Date- 30/11/2023

Parameter- TSS &amp; TDS

Test Start Date- 01/12/2023

Test End Date- 02/12/2023

Due Date- 15/12/2023

Name of Supervisor- Md. A. Rafique, Sc-'B'

Sample Code	I.W. of F.P. (gm)	F.W. of F.P. (gm)	Diff. (gm)	Sample Passed (ml)	TSS (mg/l)
2324/RDKL/WW/WW/L00132	0.0908	0.0914			
	0.0908	0.0915			
	0.0908	0.0915	0.0007	100	7
" /L00133	0.0909	0.0954			
	0.0910	0.0954			
	0.0910	0.0954	0.0044	100	44
" /L00134	0.0908	0.1021			
	0.0908	0.1021			
	0.0908	0.1021	0.0113	100	113
Sample Code	I.W. of Beaker (gm)	F.W. of Beaker (gm)	Diff. (gm)	Sample Passed (ml)	TDS (mg/l)
2324/RDKL/WW/WW/L00132	47.9971	48.0223			
	47.9970	48.0223			
	47.9970	48.0223	0.0253	100	253
" /L00133	57.3751	57.4448			
	57.3752	57.4449			
	57.3752	57.4450	0.0698	100	698
" /L00134	57.2811	57.3997			
	57.2811	57.3998			
	57.2811	57.3998	0.1187	100	1187

Ajeet Verma  
02/12/2023Md. A. Rafique  
02/12/23

Omit