

Sample Requisition No.: RDKL/EW/2324/SK00057, RDKL/EW/2324/SK00058

Parameter: PH

Name of Supervisor: Md. A. Rafique

Date of analysis: 07/02/2024

S.No.	Sample Code	PH Value	Temp.
01.	Std.	PH 4.0 $\approx$ 3.98	25°C $\pm$ 1°C
02.	Std.	PH 7.0 $\approx$ 7.1	
3.	Std.	PH 9.8 $\approx$ 9.15	
4.	2324/RDCL/EW/EW/200099	7.5	
5.	" /L00100	7.6	
6.	" /L00101	8	
7.	" /L00102	7.7	
8.	" /L00103	7.4	
9.	" /L00104	7.6	
10.	" /L00105	7.7	
11.	" /L00106	7.5	
12.	" /L00107	7.6	
13.	" /L00108	7.5	
14.	" /L00109	7.8	
15.	" /L00110	7.7	
16.	" /L00111	7.7	

By

07.02.24

Sample Requisition No.: RDKL/EW/2324/SK00057, RDKL/EW/2324/SK00058  
Parameter: EC

Name of Supervisor: Md. A. Rafique

Date of analysis: 07/02/2024

S.No.	Sample Code	EC Value	Temp.
01.	Std.	1914 $\mu$ S/cm	25°C $\pm$ 1°C
02.	2324/RDCL/EW/EW/L00099	210	
03.	" /L00100	158	
04.	" /L00101	193	
05.	" /L00102	453	
06.	" /L00103	254	
07.	" /L00104	240	
08.	" /L00105	4850	
09.	" /L00106	157	
10.	" /L00107	226	
11.	" /L00108	226	
12.	" /L00109	159	
13.	" /L00110	149	
14.	" /L00111	152	

By

By

07.02.24