

Sample Requisition No: RDKL/FN/232A/SR00063

Allocation Date: 12.02.21

Test start Date: 13.02.21

Test end Date: 13.02.21 & 14.02.21

Parameter: pH, EC, TSS & TDS

Due Date: 19.02.21

Name of supervisor: Md. A. Rafique

S.No.	Sample code	pH value	Temp (°C)
01.	Std	pH 7.0 = 3.98	(25 ± 1°C)
02.	Std	pH 7.0 = 6.97	
03.	Std	pH 9.18 = 9.17	
04.	L00126	7.2	✓

S.No.	Sample code	EC value	Temp (°C)
01.	Std	1415 µs/cm	(25 ± 1°C)
02.	L00126	1097 µs/cm	✓

S.No.	Sample code	I.Wt. of F.P (gm)	F.Wt. of F.P (gm)	Diff (gm)	Sample Panes (ml)	TSS (mg/L)
01.	L00126	0.0897	0.0910			
		0.0898	0.0910			
		0.0898	0.0910	0.0012	100	12

S.No.	Sample code	I.Wt. of Beaker (gm)	F.Wt. of Beaker (gm)	Diff (gm)	Sample Panes (ml)	TDS (mg/L)
01.	L00126	46.0073	46.0630			
		46.0074	46.0629			
		46.0074	46.0630	0.0556	100	556

Reaction Mandat.
15.02.21

Handwritten signature

Handwritten number: 160221