

Sample Requisition NO: RDKL/NW/2324/SR00004

Allocation Date: 05.10.23

Test start Date: 05.10.23

Test End Date: 05.10.23

Parameter: pH

Due Date: 19.10.23

Name of supervisor: Md. A. Rafique

S/NO.	Sample Code	pH value	Temperature
01.	std	pH 9.0 = 9.0	25°C
02.	std	pH 7.0 = 7.1	25.5°C
03.	std	pH 9.18 = 9.17	25°C
04.	2324/RDKL/NW/NW/L00004	6.83 ≈ 6.8	24.7°C
05.	" / L00005	7.6	24.7°C

Raufim Mandal.  
09.10.23

Md. A. Rafique  
19.10.23