

Sample Repquisition Date No: RDKL/WW/ 232A/SR00092

Allocation Date: 23.11.23

Test start Date: 23.11.23

Test end Date: 23.11.23

Parameter: pH

Due Date: 08.12.23

Name of supervisor: Md. A. Rafique

<u>Sl NO.</u>	<u>Sample Code</u>	<u>pH value</u>	<u>Temp. (°C)</u>
01.	std		
02.	std	pH 4.0 = 4.0	25
03.	std	pH 7.0 = 7.0	25
04.	std	pH 9.18 = 9.16	25
05.	232A/RDKL/WW/WW/L00092	7.0	25
	" / L00093	6.9	25

Raxim Mandal.

28.11.23

Guest


07.11.23

Sample Registration No: RDKL/NW/232A/GR00042

Allocation Date: 23.11.23

Test Start Date: 7.12.23

Test End Date: 8.12.23

Parameter: TSS & TDS

Due Date: 8.12.23

Name of supervisor: Md. A. Rafique

Sl. No.	Sample ID	I. wt of F.P (gm)	F. wt of F.P (gm)	Diff (gm)	Sample passed (ml)	TSS (mg/l)
01.	L00092	0.0884	0.0970			
		0.0885	0.0970			
		0.0885	0.0972	0.0086	100	86
02.	L00093	0.0876	0.0889			
		0.0876	0.0888			
		0.0876	0.0889	0.0013	100	13

Sl. No.	Sample ID	I. wt of Beaker (gm)	F. wt of Beaker (gm)	Diff (gm)	Sample passed (ml)	TDS (mg/l)
01.	L00092	47.7254	47.7647			
		47.7255	47.7648			
		47.7256	47.7648	0.0392	100	392
02.	L00093	51.7623	51.8032			
		51.7625	51.8033			
		51.7625	51.8033	0.0408	100	408

Ranjan Mandal

18.12.23

18.12.23