

Sample Requisition No: RDKL/FW/232A/SR00032

Allocation Date: 27.11.23

Test start Date: 28.11.23

Test end Date: 28.11.23

Parameter: pH & EC

Due Date: 12.12.23

Name of supervisor: Md. A. Rafique

Sl no.	Sample Code	pH value	Temp (°C)
01.	Std	pH 7.0 = 7.0	25
02.	Std	pH 7.0 = 7.0	25
03.	Std	pH 9.18 = 9.16	25
04.	232A/RDKL/FW/FW/600032	7.9	25

Sl no.	Sample Code	Ec value	Temp (°C)
01.	Std	1415 μ s/cm	25
02.	232A/RDKL/FW/FW/600032	809 μ s/cm	25

Rautin Mandat.
28.11.23

Amir

28.11.23

Sample Requisition No: RDKL/FW/232A/SR00032
 Allocation Date: 27.11.23
 Test Start Date: 7.12.23
 Test end Date: 8.12.23
 Parameter: TSS & TDS
 Due Date: 10.12.23
 Name of supervisor: Md. A. Rafique

Sample ID	I. WT of F.P (gm)	F. WT of F.P (gm)	Diff (gm)	Sample volume (ml)	TSS (mg/L)
232A/RDKL/FW/FW/L00032	0.0872	0.0873			
	0.0872	0.0875			
	0.0872	0.0875	0.0003	100	BLG

Sample ID	I. WT of Beaker (gm)	F. WT of Beaker (gm)	Diff (gm)	Sample volume (ml)	TDS (mg/L)
L00032	50.3584	50.4016			
	50.3585	50.4017			
	50.3586	50.4017	0.0431	100	431