

Sample Registration NO: RDKL/FW/2324/SR00017

Parameters: pH, Conductivity

Allotment Date: 13/10/2023

No. of Samples: 01

Date of Sample Analysis: ~~13/10/2023~~ 13/10/2023

Date of Result Submission: 17/10/2023

Name of the Supervisor: Md. Abdur Raufique, Sec-B

S.No.	Sample Code	pH	Conductivity μs/cm or μmhos/cm	Temperature (in °C)
01.	pH Buffer 4.0	3.9412 ≈ 3.9	—	25
02.	pH Buffer 7.0	7.0222 ≈ 7.0	—	25
03.	pH Buffer 9.18	9.1252 ≈ 9.13	—	25
04.	Standard KCl 0.01(M)	—	1414	25
05.	2324/RDKL/FW/FW/L00013	7.4122 ≈ 7.41	974	25

[Signature]
17/10/2023

Dr. Raufique
20.10.23