

Sample Requisition - RDKL/WW/2324/SR00094
 RDKL/WW/2324/SR00096
 RDKL/WW/2324/SR00097

Test Parameter - Total Nitrogen


Date of Analysis - 29/11/24

Test Material - Water

Name of Supervisor - Mr. Md. A. Refique

Requisition No.	Sample Code	Sample Vol. (ml)	Absorbance	Dilution	Total Nitrogen (mg/L)
RDKL/WW/2324/SR00094	2324/RDKL/WW/WW/L00304	25	0.296	25	20
	2324/RDKL/WW/WW/L00305	25	0.308	25	21
	2324/RDKL/WW/WW/L00311	25	0.243	25	16
	2324/RDKL/WW/WW/L00312	25	0.371	25	25
	2324/RDKL/WW/WW/L00313	25	0.327	25	22
RDKL/WW/2324/SR00096	2324/RDKL/WW/WW/L00314	25	0.371	25	25
	2324/RDKL/WW/WW/L00315	25	0.36	25	24
	2324/RDKL/WW/WW/L00316	25	0.315	50	42
	2324/RDKL/WW/WW/L00317	25	0.303	50	41
	2324/RDKL/WW/WW/L00318	25	0.346	50	47
	2324/RDKL/WW/WW/L00319	25	0.274	50	37
	2324/RDKL/WW/WW/L00320	25	0.39	50	53
	2324/RDKL/WW/WW/L00321	25	0.236	50	32
	2324/RDKL/WW/WW/L00323	25	0.325	50	44
	2324/RDKL/WW/WW/L00324	25	0.417	50	56
	2324/RDKL/WW/WW/L00325	25	0.288	50	39
	2324/RDKL/WW/WW/L00326	25	0.3	50	40
	2324/RDKL/WW/WW/L00327	25	0.321	50	43
	2324/RDKL/WW/WW/L00328	25	0.292	50	39
RDKL/WW/2324/SR00097	2324/RDKL/WW/WW/L00303	25	0.440	50	59

Concentration (mg/L)	Absorbance
0	0
0.05	0.01
0.1	0.022
0.25	0.09
0.5	0.173
0.75	0.269
1	0.387
Slope	0.3711



Handwritten signature/initials

Handwritten signature/initials